Adverse Events Associated With Robotic-Assisted Spine Surgery: An Analysis of the U.S. Food and Drug Administration MAUDE Database

Andrew Moon, Nicholas Joseph Coccoluto, Nicholas R Pagani¹, Michael A Moverman¹, Richard N Puzzitiello, Mariano Menendez, Brian Kwon

¹Tufts Medical Center

INTRODUCTION:

Despite the purported benefits of robotic assistance in spine surgery, there remains a risk of associated adverse events leading to patient morbidity and surgical delay. The aim of this study was to determine the nature and frequency of adverse events associated with robotic-assisted spine surgery. METHODS:

All adverse event reports to the FDA MAUDE database involving robotic-assisted spine surgery were reviewed from January 2018 to December 2021.

RESULTS:

A total of 236 adverse event reports were included. The most frequently reported type of adverse event was unexpected robotic arm movement (11/236, 4.7%). There were 161 reports of surgical delay with delays ranging from less than an hour up to over three hours. There were 108 cases that required conversion to manual surgery. A total of 52 patient injuries were reported, 22 of which required surgical re-intervention. Of these, spinal canal breech (18/236, 7.6%) was the most common. Additional injuries reported included nerve injury with possible associated pain, muscle weakness and sensory changes, infection, tissue damage, implant failure and cyst formation.

DISCUSSION AND CONCLUSION:

As the adoption of robotic technology in spine surgery continues to expand in the United States, improved understanding of associated adverse events can help inform both patients and surgeons. Increased awareness of the potential limitations of this technology may help prevent future adverse events with the goal of improving patient outcomes.

		-			Table 5 Primary Types of Reported Patient Injuri		
Table 1 Adverse Events Reported to the FDA Maude Database		Components	Total				Requiring
Total Adverse Events	236		Iotai	Porcontago		Total	Surgical
	250		254*	reicentage		n (<u>%)</u> a	Re-Intervention
Component Type Involved, n (%)		Inaccurate Guidance Causing Deviated Trajectory	101	39.8%			n (%) b
Software	116 (49.2)	Inaccurate Instrument Tracking	18	7.1%		All 59*	
Mechanical	61 (25.8)	Soft Tissue Pressure Causing Deviated Trajectory	10	3.9%	Spinal Canal Breech	18 (34.6)	8 (44.4)
D-th	10 (7 C)	Failed Surgical Arm Accuracy Test	9	3.5%	Pain	13 (25)	10 (76.9)
Both	18 (7.6)	Registration Difficulty	8	3.1%	Weakness	9 (17.3)	7 (77.8)
Other	41 (17.4)	Software Crash	7	2.8%	Infection	5 (9.6)	0 (0)
Surgical Delay Reported n (%)	162 (68 6)	Anatomy Shift	6	2.4%	Unspecified Nerve Damage	5 (9.6)	0 (0)
1 FO minutes	102 (00.0)	Unprompted Surgical Arm Movement	6	2.4%	Tissue Damage	2 (3.8)	0 (0)
1-59 minutes	136 (57.6)	Surgical Arm Component Loose	5	2.0%	Sensory Changes	1 (1.9)	1 (100)
>60 minutes	23 (9.7)	Camera Connection Failure	4	1.6%	Implant failure	1 (1.9)	1 (100)
Not Specified	3 (1.3)	Surgical Arm Joint Malfunction	4	1.6%	Cyst Formation	1 (1.9)	1 (100)
		Other Surgical Arm Manufaction	4	1.076	Unknown	4 (7.7)	0 (0)
Conversion to Manual, n (%)	97 (41.1)	3Define scan Issue	4	1.0%	a Percentage based on total patients who were i	niured (52)	
Case Aborted, n (%)	8 (3.4)	Controller Communication Error	2	0.8%	b Percentage based on individual patient injury	, , ,	
Patient Injury, n (%)	51 (21.6)	Surgical Arm Vibration	2	0.8%	*Some injuries fell under multiple classifications		
Surgical Ba Intervention due to Injuny n (%)	22 (0.2)	Instrument Stuck in Place	2	0.8%			
Surgical Re-Intervention due to Injury, IT (%)	22 (9.5)	Drill Vibrations	2	0.8%			
Total Surgical Re-Intervention, n (%)	37 (15.7)	3D Marker Crossthreading	1	0.4%			
MAUDE, Manufacturer and User Facility Device Experience; FDA, US		Accuracy Pointer Bent	1	0.4%			
Food and Drug Administration		C Arm Calibration Issue	1	0.4%			
- Demonstrans based on total advance succets		Dilator Attachement Broken	1	0.4%			
a Percentage based on total adverse events		Display Screen	1	0.4%			
		Draping Issue	1	0.4%			
		Excessive Drilling Force	1	0.4%			
		O Arm Connection Failure	1	0.4%			
		Stealth Camera Failure	1	0.4%			
		Unprompted Emergency Stop	1	0.4%			
		Cable Connection Failure	1	0.4%			
		Dull Drill Bit	1	0.4%			
		Navlock Tracker Malfunction	1	0.4%			
		Instrument Verification Problem	1	0.4%			
		Segmentation Failure	1	0.4%			

Percentage based on total adverse events for each surgery 18 of the 236 reports had both a mechanical and software component